



中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7105157152
Report No: ZFLJ2644811	Page 1 of 3



Applicant Information

Applicant Name : SANKOM SA
Applicant Address : SANKOM Building Rue de la Moliere 2 CH-2800 Delemont

Sample Information

Sample Description : AVEVITTA PROTECT 2.0
Brand : AVEVITTA
Sample Quantity : 21 pieces
Model : AV-020

- Sample Receiving Date : 2020-08-28
- Report Date : 2020-08-30
- The original sample is stucked on the last paper.

Test Performed

Judgement according to:
ASTM F2100-2019e1 Standard Specification for Performance of Materials Used in Medical Face Masks
Level 3
- Selected test(s) as requested by applicants. For details, refer to attached page(s).

Pronounce

The results shown in this report refer only to sample(s) tested unless otherwise stated.
All the items are performed in the standard conditions, except the noted cases.
Except for the requirement of the client, the test results and the conformity judgement of this report do not take the uncertainty of the test results into account.

Signed for and on behalf of
CNTAC Testing Service Co.,Ltd.(Foshan)

Approved by

张志荣





中国认可
国际互认
检测
TESTING
CNAS L1842



佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7105157152
Report No: ZFLJ2644811	Page 2 of 3



--- Test Result ---

Test Result	Requirements	Judgement
1. Flammability 16 CFR part 1610-2019		
Unit: <grade>		
Result	Class 1	Class 1
Combustion regime	DNI	Pass
Remark:DNI = Did not ignite.		





中国认可
国际互认
检测
TESTING
CNAS L1842




佛山中纺联检验技术服务有限公司
CNTAC Testing Service Co.,Ltd.(Foshan)

Testing Report	Security website: www.fcl-sz.org.cn Security code: 7105157152
Report No: ZFLJ2644811	Page 3 of 3



Original Sample



===== End of Report =====

